Accepted Manuscript

Title: Development of a simple analytical method for the simultaneous determination of paracetamol, paracetamol-glucuronide and *p*-aminophenol in river water



Author: Lúcia H.M.L.M. Santos Paula Paíga Alberto N. Araújo Angelina Pena Cristina Delerue-Matos M. Conceição B.S.M. Montenegro

PII:	S1570-0232(13)00243-2
DOI:	http://dx.doi.org/doi:10.1016/j.jchromb.2013.04.032
Reference:	CHROMB 18384
To appear in:	Journal of Chromatography B
Received date:	2-11-2012
Revised date:	16-4-2013
Accepted date:	23-4-2013

Please cite this article as: L.H.M.L.M. Santos, P. Paíga, A.N. Araújo, A. Pena, C. Delerue-Matos, M.C.B.S.M. Montenegro, Development of a simple analytical method for the simultaneous determination of paracetamol, paracetamol-glucuronide and *p*-aminophenol in river water, *Journal of Chromatography B* (2013), http://dx.doi.org/10.1016/j.jchromb.2013.04.032

This is a PDF file of an unedited manuscript that has been accepted for publication. As a service to our customers we are providing this early version of the manuscript. The manuscript will undergo copyediting, typesetting, and review of the resulting proof before it is published in its final form. Please note that during the production process errors may be discovered which could affect the content, and all legal disclaimers that apply to the journal pertain.

	1	
1 2 3	2	Development of a simple analytical method for the simultaneous determination
4 5	3	of paracetamol, paracetamol-glucuronide and p -aminophenol in river water
6 7 8	4	
9 10 11	5	
12 13	6	Lúcia H.M.L.M. Santos ^a , Paula Paíga ^b , Alberto N. Araújo ^a , Angelina Pena ^c , Cristina
14 15 16	7	Delerue-Matos ^b , M. Conceição B.S.M. Montenegro ^{a*}
17 18	8	
19 20 21	9	
22	10	
23 24 25	11	^a REQUIMTE, Faculty of Pharmacy, University of Porto, Rua Jorge Viterbo
26 27 28	12	Ferreira 228, 4050-313 Porto, Portugal
29 30	13	^b REQUIMTE, Instituto Superior de Engenharia do Porto, Instituto Politécnico do
31 32 33	14	Porto, Rua Dr. António Bernardino de Almeida, 431, 4200-072 Porto, Portugal
34 35	15	^c Group of Health Surveillance, Center of Pharmaceutical Studies, University of
36 37 28	16	Coimbra, Health Sciences Campus, Azinhaga de Santa Comba, 3000-548 Coimbra,
38 39 40	17	Portugal
41 42 43	18	
44 45	19	*Corresponding author: Phone: +351 220428677; fax: +351 226093390; e-mail:
46 47 48	20	<u>mcbranco@ff.up.pt</u>
49 50 51 52	21	
53		
54 55		
56		
57		
58 59		
60		
61 62		
63		1
64		
65		

	1	
	1 2	
	2	
	3 4 5	
	4 5	
	5	
	6 7 8	
	/	
	8 9	
	0	
1	1	
L	2	
L	3	
L	4	
L	5	
L	6	
L	./	
L	8	
L	9	
2	0	
2	234567890123456789012345678	
2	2	
2	3	
2	4	
2	5	
2	6	
2	7	
2	8	
2	9	
3	0	
3	1	
3	2	
3	3	
3	4	
3	5	
3	6	
3	7	
3	8	
3	9	
	0	
1		
1	2	
1	3	
1	4	
1	5	
1	6	
1	7	
1	8	
1	9	
5	0 1 2 3 4 5 6	
5	1	
5	2	
5	3	
5	4	
5	5	
5	6	
5	7	
5	8	
5	9	
5	7 8 9 0 1 2 3 4	
5	1	
5	2	
5	3	
5	4	

Abstract

22

23

24 Paracetamol is among the most worldwide consumed pharmaceuticals. Although its 25 occurrence in the environment is well documented, data about the presence of its 26 metabolites and transformation products is very scarce. The present work describes the 27 development of an analytical method for the simultaneous determination of paracetamol, its principal metabolite (paracetamol-glucuronide) and its main 28 29 transformation product (*p*-aminophenol) based on solid phase extraction (SPE) and high 30 performance liquid chromatography coupled to diode array detection (HPLC-DAD). 31 The method was applied to analysis of river waters, showing to be suitable to be used in 32 routine analysis. Different SPE sorbents were compared and the use of two Oasis WAX 33 cartridges in tandem proved to be the most adequate approach for sample clean up and 34 pre-concentration. Under optimized conditions, limits of detection in the range of 40 to 35 67 ng/L were obtained, as well as mean recoveries between 60 and 110% with relative 36 standard deviations (RSD) below 6%. Finally, the developed SPE-HPLC/DAD method 37 was successfully applied to the analysis of the selected compounds in samples from 38 seven rivers located in the north of Portugal. Nevertheless all the compounds were 39 detected, it was the first time that paracetamol-glucuronide was found in river water at 40 concentrations up to $3.57 \,\mu g/L$.

Keywords: Paracetamol, paracetamol-glucuronide, *p*-aminophenol, solid phase
extraction, HPLC, river water

46

65

41

42

1. Introduction

Paracetamol (acetaminophen or N-acetyl-4-aminophenol) is one of the most popular and widely used medicines for the treatment of pain and fever, both as an over-the-counter (OTC) and as a prescribed medicine. It can be used in a wide range of patients, including children, pregnant women or the elderly. Following oral administration, approximately 90% of paracetamol is metabolized, being conjugated with glucuronide (40–67%) and, in a less extent, with sulphate (20–46%), to form inactive metabolites, which are eliminated in urine together with a small fraction of unchanged paracetamol (<5%) [1]. Although paracetamol presents a high removal efficiency (approximately 99%) in WWTPs [2, 3], it has been detected in their effluents at concentrations up to low microgram per litre [4-6], contributing to its entrance into surface waters [7-9]. Once in the environment, paracetamol is mainly degraded by microorganisms, which are capable of using it as carbon and energy sources [10]. Although paracetamol is not highly persistent in the environment, continuous input overrules its high transformation rate [11], thus, it can adversely affect aquatic organisms. Acute toxicity effects in the invertebrate Daphnia magna (EC₅₀ ranging from 26.6 to 50 mg/L) [12-14], the marine bacterium Vibrio fischeri (EC₅₀ = 549.7 mg/L) and the fish *Oryzias latipes* (EC₅₀ = >160 mg/L) [12] have been reported. Effects on cell cultures with EC_{50} values of 19 mg/L have also been described [13]. Nowadays analytical methodologies described in literature are mainly focused in multi-residues methods that allow the simultaneously determination of paracetamol together with a large number of pharmaceuticals from several therapeutic groups [15-17]. Most of them are principally focused in parent compounds and rarely analyze metabolites and/or transformation products. At present, methods reported for the determination of paracetamol metabolites are focused in biological matrices [18-20]

rather than in environmental ones [21]. On the other hand, paracetamol may also be degraded, both during wastewater treatment and in the environment, giving different transformation products [22, 23]. p-Aminophenol was identified as its main transformation product, and its presence in wastewater samples was reported [23]. However the origin of *p*-aminophenol cannot only be attributed to the degradation of paracetamol, since it is also widely used in industrial applications and is known as a transformation product from pesticides. Furthermore, *p*-aminophenol was also described as the primary degradation product of paracetamol during the storage of its medicinal formulations [24].

High-performance liquid chromatography (HPLC) coupled to tandem mass
spectrometry (MS/MS) has been designated as technique of choice for the determination
and quantification of pharmaceuticals in environmental samples [25]. However, these
equipments are still very expensive and they are not available in many laboratories for
routine analysis. On the other hand, almost all laboratories have HPLC systems with
diode array UV absorbance and/or fluorescence detection that may effectively be used
for the analysis of pharmaceuticals in environmental samples [26-29].

Due to the complexity of environmental samples, analysis of pharmaceuticals has to be preceded by a pre-concentration step, which allows the detection of low concentrations and simultaneously removes the interferences. This is often performed by solid phase extraction (SPE). Generally, polymeric sorbents, like Oasis HLB, are the most used for pre-concentration of pharmaceuticals from aqueous matrices [16, 17, 30-32], though mixed-mode ion-exchange sorbents have also been described [33, 34]. Thus, the present work describes the development and validation of an analytical method based on off-line SPE, using a mixed mode reversed phase/ anionic exchange sorbent, followed by LC-DAD for the determination of paracetamol, its main metabolite

(paracetamol-glucuronide) and its principal transformation product (*p*-aminophenol) in river waters. The performance and application of this method is important, since allows the simultaneous monitoring of parent compound, metabolite and transformation product as well as the evaluation of their environmental interdependence, using one of the most worldwide consumed pharmaceuticals (paracetamol) as example. Finally, the developed methodology was successfully applied to the analysis of the selected compounds in seven rivers from north of Portugal. To our knowledge this is the first time that paracetamol-glucuronide was found in surface waters.

2. Materials and methods

2.1. Chemicals and reagents

Paracetamol (PCT) (acetaminophen), p-aminophenol (PAP) (4-aminophenol) and paracetamol-glucuronide (PCT-G) (ρ -acetamidophenyl β -D-glucoronide) sodium salt were purchased from Sigma-Aldrich (Steinheim, Germany). All standards were of high purity grade (>93%). HPLC-grade methanol, HPLC-grade acetonitrile and HPLC-grade acetone, n-hexane and formic acid (purity $\geq 98\%$) were obtained from Merck (Darmstadt, Germany), hydrochloric acid 37% and glacial acetic acid (purity $\geq 99.7\%$) were purchased from Carlo Erba (Rodano, Italy), ammonia 25% was obtained from Panreac (Barcelona, Spain), ammonium hydroxide solution, ammonium acetate (purity \geq 98%), ethyl acetate and dichloromethane were purchased from Sigma-Aldrich (Steinhein, Germany). HPLC-grade water (18.2 M Ω cm) was obtained by purifying deionised water in a Milli-Q Simplicity 185 system (Millipore, Molsheim, France). Individual stock standard solutions were prepared for each compound by dissolving 10 mg of powder in 10 mL of methanol, obtaining a final concentration of

1000 mg/L, and stored at -20 °C. Stock standard solutions were renewed every week. An intermediate standard solution was daily prepared by mixing the three individual stock solutions and diluting with a mixture methanol-water (10:90, v/v) to give a final concentration of 10 mg/L and kept at 4 °C. Working standard solutions were also prepared in a mixture methanol-water (10:90, v/v) by dilution of appropriate amounts of the intermediate solution. Amber glassware was used to prevent light degradation. These working standard solutions were used for preparation of the calibration curve and for spiking samples in the validation study.

All standard solutions and sample extracts were filtered through a 0.20 µm
PTFE syringe filter (Teknokroma, Barcelona, Spain) and homogenised using a vortex
mixer (VWR, Radnor, Delaware, USA). All chromatographic solvents were filtered
through a 0.20 µm nylon membrane filter (Supelco, Bellefonte, PA, USA) using a
vacuum pump (Dinko D-95, Barcelona, Spain) and degassed for 15 min in an ultrasonic
bath (Raypa[®] Trade, Terrassa, Spain).

136 SPE cartridges used were Oasis[®] MAX (60 mg, 3 mL), Oasis[®] WAX (150 mg, 6 137 mL), Oasis[®] MCX (150 mg, 6 mL) and Oasis[®] HLB (200 mg, 6 mL) from Waters 138 (Mildford, MA, USA), LiChrolut[®] EN/RP-18 (EN 40-120 μ m, 100 mg (bottom) and 139 RP-18 40-63 μ m, 200 mg (top), 6 ml) from Merck (Poland), StrataTM-SDB-L (500 mg, 140 6 mL) and StrataTM-X (200 mg, 3 mL) from Phenomenex (USA), and Enviro-clean[®] (C₈ 141 and quaternary amine, 1000 mg, 6 mL) from Unit Chemical Technologies (UCT), Inc. 142 (Bristol, PA, USA).

2.2. Sample collection

River water (2.5 L) was collected from seven rivers located in the north of Portugal,
which is one of the most densely populated areas of the country. Sample collection,

	1	
	2	
	2	
	5	
	4	
	5	
	6	
	7	
	, Q	
	0	
_	9	
1	0	
1	1	
1	2	
1	3	
1	1	
1	т г	
T	5	
T	6	
1	7	
1	8	
1	9	
2	0	
2	1	
2	Ť	
2	2	
2	3	
2	4	
2	5	
2	6	
າ 2	3456789012345678901234567890123456789012345678	
2	/	
2	8	
2	9	
3	0	
3	1	
2	2	
2	2	
3	3	
3	4	
3	5	
3	6	
3	7	
3	8	
2	9	
2		
	0	
4		
4	2	
4		
	4	
	5	
4	s c	
4	0	
4		
4		
4	9	
5	0	
5	1	
55	2	
	-)	
5 5	د	
5	4	
5	5	
5	6	
5	7	
5	8	
5	q	
6		
6		
6	2	
6	3	
6	4	
6	5	

preservation and storage were done according to the US EPA Method Guideline [35].
River samples were collected on the river side in amber glass bottles and kept
refrigerated (± 4°C) during the transport to the laboratory. Samples were collected along
one week in September 2011.

- 151
- 152

2.3. Sample pre-treatment and extraction

153 River water samples were vacuum filtered through 1.2 μm glass microfiber filters

154 (GF/C, Whatman, UK), followed by 0.20 µm nylon membrane filters (Supelco,

155 Bellefonte, PA, USA) and stored at -20 °C, until extraction.

156 For the SPE procedure a vacuum manifold system (Phenomenex, USA) was 157 used. Two Oasis WAX cartridges were initially conditioned, in separate, with 2 mL of 158 methanol, 2 mL of HPLC-grade water, and 2 mL of HPLC-grade water pH 7 (pH 159 adjusted with ammonia) at a flow rate of 1 mL/min. After that, the SPE cartridges were 160 connected in tandem and 50 mL of river water (pH adjusted to 7 with ammonia) were 161 loaded onto the cartridges at a flow rate of 1 mL/min. Finally, analytes were eluted with 162 5 mL of methanol and 5 mL of 5% ammonium hydroxide in methanol at a flow rate of 1 163 mL/min, and the eluates were pooled in one single collection vial. Extracts were 164 evaporated to dryness under a gentle stream of nitrogen and reconstituted in 250 µL of a 165 mixture methanol-water (10:90, v/v), allowing a pre-concentration factor of 200.

- 166
- 167

2.4. Liquid chromatography

168 Chromatographic analysis was performed on a Nexera Ultra-High Performance 169 Liquid Chromatography system (Shimadzu Corporation, Kyoto, Japan) equipped with 170 two solvent delivery modules LC-30 AD, a column oven CTO-20 AC, an autosampler 171 SIL-30 AC and an UV/Vis photodiode array detector SPD-M20A. The system was

controlled by a system controller CBM-20A. Two chromatographic systems were used in this work in order to evaluate the performance of the method for HPLC and fast-HPLC analysis. On the former, separation was carried out using a Luna C18(2) column (150 x 4.6 mm i.d., 5 µm particle size) (Phenomenex, USA) that was kept at 25 °C, while sample vials were kept at 4 °C. An injection volume of 40 µL was used. The optimized mobile phase consisted of 10 mM ammonium acetate/acetic acid (pH 6) as solvent A and acetonitrile as solvent B, using a flow rate of 1 mL/min. The gradient elution was performed as follow: 0-9.0 min, 97-74.8% A; 9.0-10.0 min, return to initial conditions; 10.0–15.0 min equilibration of the column. For fast-HPLC analysis, a Shim-pack XR-ODS (100 x 4.6 mm i.d., 2.2 µm particle size) (Shimadzu, Kyoto, Japan) column was used and it was kept at 25 °C. The separation was performed using the same mobile phase composition (10 mM ammonium acetate/acetic acid (pH 6) as solvent A and acetonitrile as solvent B) at a flow rate of 2 mL/min. The gradient elution was: 0-1.0 min, 96-91.6% A; 1.0-1.5 min, 91.6-80% A; 1.5-3.0 min, 80% A; 3.0-3.5 min, returned to initial conditions; 3.5-4.0 min, equilibration of the column. An injection volume of 10 µL was used and sample vials were kept at 4 °C. The analytes were monitored at their maximum UV wavelengths, namely 243, 232 and 245 nm for paracetamol-glucuronide, p-aminophenol and paracetamol, respectively. Lab Solutions software (Shimadzu Corporation, Kyoto, Japan) was used for control and data processing.

2.5. Method validation

Method validation was performed for both HPLC and fast-HPLC analysis, in order to evaluate which chromatographic technique might be more suitable for application to real samples.

Target analytes were identified in the chromatograms by comparison of the
retention time of the peaks obtained with these ones of a standard solution.
Simultaneously the identification of the analytes was also confirmed comparing the
corresponding UV spectra of the peaks of the sample and of standard solution
chromatograms. For the quantification of the analytes in real samples the standard
addition method was used.

The linearity of the method was established by setting calibration curves using linear regression analysis over the concentration range of 50 to 1500 μ g/L; however final concentrations tested depend of the sensitivity reached for each analyte. Method detection limit (MDL) and method quantification limit (MQL) were determined as the minimum amount detectable of analyte with a signal-to-noise ratio of 3 and 10, respectively.

Method accuracy (expressed as recovery percentage) and precision (expressed in terms of relative standard deviation (RSD)) were evaluated by recovery studies of the analytes in river water, spiked at different final concentration levels (0.75, 2.50, 3.75 and 5.00 μ g/L). Experiments were performed in triplicate (n = 3). Recoveries were determined comparing the concentrations obtained with the initial spiking levels. As river water may contain the target analytes, blanks (samples without standard solution addition) were analyzed in order to determine their concentrations, which were then subtracted to the spiked river water.

217 Method precision was determined by repeated intra- and inter-day analysis, through 218 five successive injections of a river water sample spiked with a standard mixture 219 containing all the analytes at a final concentration of $5.00 \mu g/L$ in one day and its 220 injection in five consecutive days, respectively.

221	The influence of the matrix in the UV signal was evaluated by preparing a
222	calibration curve in river extract, and comparing it with one achieved for the standards
223	prepared in a mixture of methanol-water (10:90, v/v). A blank (sample with no addition
224	of the standards) was simultaneously assayed in order to subtract the levels of the target
225	analytes present in the sample.
226	
227	
228	3. Results and discussion
229	3.1. Solid phase extraction optimization
230	The SPE method was optimized using ultra-pure and river water.
231	The selection of the most adequate SPE sorbent is a critical point in the
232	development of an extraction procedure; therefore the performance of different SPE
233	cartridges was evaluated and the recoveries in ultra-pure water, pH adjusted to 7, are
234	shown in Figure 1. As can be seen, paracetamol could be efficiently recovered by
235	almost all SPE cartridges (with the exception of Enviro-clean [®]), while for its metabolite
236	higher recoveries (72–103%) were obtained with the mixed mode reversed
237	phase/anionic exchange sorbents Oasis MAX and WAX, since at pH 7, paracetamol-
238	glucuronide is negatively charged (Figure S1, Supplementary data), establishing ionic
239	interactions with those sorbents. Therefore, polymeric sorbents such as Oasis HLB,
240	Strata-X and Strata SDB-L, which extract analytes by reversed phase mechanisms, were
241	not able to recover this compound. Furthermore, <i>p</i> -aminophenol showed the lowest
242	recovery, which did not exceed 43%, using the Oasis WAX cartridges, and it was not
243	recovered at all with Enviro-clean [®] , given that at pH 7, p -aminophenol is in a neutral
244	form (Figure S1, Supplementary data), and this mixed mode sorbent (cationic exchange
245	and C_8) is not able to interact with <i>p</i> -aminophenol by reversed phase mechanisms. As it

1 2 3	
4 5 7 8 9 10	
9 10 11 12	
13 14 15 16	
17 18 19 20	
$\begin{array}{c} 1\\ 1\\ 1\\ 1\\ 3\\ 1\\ 1\\ 1\\ 1\\ 1\\ 1\\ 1\\ 1\\ 1\\ 1\\ 1\\ 1\\ 1\\$	
25 26 27 28 29	
30 31 32 33	
34 35 36 37	
40 41	
42 43 44 45 46	
47 48 49 50	
51 52 53 54	
55 56 57 58	
59 60 61 62	
63 64 65	

251

252

253

is depicted in Figure 1, the hydrophilic polymeric sorbents Oasis HLB and Strata X
yielded similar recoveries for all analytes, which is in agreement with their similar
physico-chemical properties [36]. Comparing the different SPE sorbents tested, it can be
seen that highest average recoveries were achieved for all the analytes with Oasis WAX
(Figure 1).

Insert Figure 1

254 Mixed-mode ion-exchange sorbents allow two types of interaction mechanisms, namely reversed phase and ionic-exchange. Thus, taking special attention to pH and 255 256 solvents employed in each step of SPE protocol, analytes can be selectively eluted [37]. 257 In this context, the effect of sample's pH was studied for the mixed-mode ion-exchange 258 sorbents (Oasis MAX, WAX and MCX) in order to enhance the *p*-aminophenol 259 recovery. Figure 2 shows the recoveries obtained in ultra-pure water. For paracetamol-260 glucuronide, better recoveries were achieved using reversed phase/anionic exchange 261 sorbents (Oasis MAX and WAX) at a pH range between 3 and 7, which is in agreement 262 with its pKa (Table 1). At this pH range paracetamol-glucuronide is negatively charged 263 (Figure S1, Supplementary data), being able to bind to the sorbent by ionic interactions, 264 since Oasis MAX and WAX are a strong and a weak anionic-exchange polymeric 265 sorbent, respectively, that are based on Oasis HLB (poly(N-vinylpyrrolidone-266 divinylbenzene) copolymer) chemically modified with quaternary amine groups 267 (dimethylbutylamine) and piperazine groups, respectively [37]. On the other hand, 268 Oasis MCX is a strong cation-exchange polymeric sorbent that has an Oasis HLB 269 skeleton chemically modified with sulfonic groups [37], therefore is not able to interact

	1
	1
	2
	3
	4
	5
	6 7 8
	7
	8
	9
1	0
1	1
1	ナ つ
1	⊿ ว
1	3
1	4
L	5
L	6
L	7
1	1234567890123456789012345678
1	9
2	0
2	1
2	2
2	3
2	4
2	5
2	5
ב כ	0 7
2 7	0
2 ר	0
2	9
3	0
3	1
3	2
3	3
3	4
3	5
3	б
3	7
3	8
3	9
4	0
1	
1	2
1	3
	4
	4 5
	5 6
	6
4	/
1	8
1	9
5	0
5	1
5	2
5	3
5	4
5	5
5	6
5	7
5	8
5	0 1 2 3 4 5 6 7 8 9
5	0
5	1
55	1 2
55	∠ 3
5 5	3 4
-	4

65

272

273

274

with a negatively charged compound, and for that reason the recoveries for Oasis MCX are very low (<20%).

Insert Table 1 and Figure 2

275 Like paracetamol-glucuronide, the highest recoveries of *p*-aminophenol were 276 also achieved with a reversed-phase/anionic exchange sorbent, namely Oasis WAX, at a 277 pH range between 6 and 8 (Figure 2b). According to *p*-aminophenol pKa, at these pH 278 values the molecule is in a neutral form (Figure S1, Supplementary data), being 279 extracted through reversed phase mechanisms. However, recoveries of *p*-aminophenol 280 using Oasis MCX were similar to that one obtained with Oasis WAX for a pH up to 6. 281 At this pH range *p*-aminophenol presents the amine group positively charged (Figure 282 S1, Supplementary data) and, therefore, is able to establish ionic interactions with the 283 sulfonic groups of the sorbent.

Relatively to paracetamol, all the mixed-mode ion-exchange sorbents tested provided good recoveries (higher than 70%) in all the studied pH range (Figure 2c), since it is in a neutral form (Figure S1, Supplementary data), being extracted through reversed phase mechanisms. Based on the results obtained, a sample's pH of 7 was chosen for further studies.

In contrast to the excellent recoveries of paracetamol and its metabolite obtained with Oasis WAX, the recovery of *p*-aminophenol may be improved using two SPE cartridges in tandem. Different possible combinations were tested using Oasis WAX, MCX and HLB. Oasis HLB was used in this study due to have been proved that, at pH 7, *p*-aminophenol is extracted by reversed phase mechanisms, so an increase in the mass sorbent of the lipophilic/hydrophilic balance Oasis HLB sorbent might improve the

recovery of *p*-aminophenol. Results showed that two Oasis WAX columns in tandem
allowed increasing the recovery of *p*-aminophenol (64%), maintaining good recoveries
of the others analytes (Table S1, Supplementary data).

Moreover, different organic solvents were tested namely methanol, acetonitrile, acetone, ethyl acetate, *n*-hexane and dichloromethane, as well as different proportions of ammonium hydroxide in methanol. No improvement on *p*-aminophenol recovery was observed comparatively to the elution protocol suggested by the manufacturer that is methanol, for the extraction of neutral compounds extracted by reversed phase mechanisms, and 5% ammonium hydroxide in methanol, to revert the ionic interactions and release the compounds extracted trough anionic exchange mechanisms. After the SPE protocol have been settle, different sample volumes were studied using two Oasis WAX cartridges in tandem. For paracetamol and paracetamol-

307 glucuronide, recoveries around 100% were achieved for sample volumes up to 100 mL,

308 although the recovery of *p*-aminophenol decreased with increasing sample volume

309 (Figure S2, Supplementary data). Thus, attending that the main goal of this work was to

310 develop an extraction procedure for the determination of paracetamol, its main

311 metabolite (paracetamol-glucuronide) and its main transformation product (*p*-

aminophenol) in river waters, a sample volume of 50 mL was chosen as a commitmentbetween the recovery of the analytes and the intended application of the developed

method.

3.2. Optimization of chromatographic conditions

317 In order to optimize the chromatographic separation, different mobile phases were 318 tested, using methanol, acetonitrile or a mixture of methanol and acetonitrile as organic 319 solvent and water with different additives, such as ammonium acetate and acetic acid at

different concentrations as aqueous phase. The best separation was achieved using 10
mM ammonium acetate/acetic acid (pH 6) and acetonitrile. This mobile phase was used
for both developed chromatographic procedures.

3.2.1. HPLC analysis

After the mobile phase composition had been established, the elution gradient and flow rate were adjusted in order to improve the chromatographic resolution as well as the peaks shape and to get the shorter analysis time. The optimum flow rate was set at 1.0 mL/min and the elution was performed on a 15 minutes gradient. Different column temperatures were also studied (23 °C, 25 °C and 30 °C) and for all the analytes the best peak shape and resolution were obtained using a temperature of 25 °C. Finally, the injection volume was tested (20, 30 and 40 μ L) and the chromatographic response was improved, without loss of resolution or peak shape, when an injection volume of 40 μ L was used. A representative chromatogram of a 10 μ g/mL standard mixture of the compounds analyzed is presented in Figure 3a.

Insert Figure 3

3.2.2. Fast-HPLC analysis

Upon reducing the particle size from chromatographic column from 5 µm to 2.2 µm,
it was possible to use higher flow rates, obtaining a faster separation. Parameters such
as elution gradient and flow rate were optimized in order to find the best
chromatographic resolution and to obtain narrower peaks in the shortest analysis time.
This was achieved using a gradient elution performed on 4 minutes with an optimal
flow rate of 2.0 mL/min. The influence of different temperatures in the range of 25 °C to

1	
2	
3 4 5 6 7 8 9	
5 6	
7 8	
10	
11	
13	
14 15	
16 17	
18 19	
20	
22	
23 24	
25 26	
27 28	
29	
31	
32 33	
34 35	
112 1314 1516 1718 2022 2232 2252 22930 312 32333 3536 3738	
38 39	
40 41	
42	
43 44	
45 46	
46 47 48	
49 50	
51	
53	
54 55	
56 57	
58 59	
60 61	
62	
63 64 65	
00	

345 35 °C was also studied and the best peak shapes was obtained for a temperature of 25 346 °C. Lastly, the injection volume was varied from 5 to 10 μ L in order to improve the 347 chromatographic response without loss of the peak shapes. This was achieved for an 348 injection volume of 10 μ L. An example of a chromatogram obtained with the optimized 349 chromatographic conditions is depicted in Figure 3b.

3.3. Method validation

The performance of the developed methods was validated in terms of sensitivity, linearity, recoveries, precision (intra- and inter-day) and interference of the matrix,

using river water. Detailed analytical quality assurance data is shown in Table 2.

Insert Table 2

357

355

356

350

351

As it is depicted in Table 2a, HPLC and fast-HPLC methods showed differences in 358 359 what concern to sensitivity, being the lowest MDL and MQL obtained for the former. 360 Linearity was studied in the range of 50 to 1500 μ g/L, setting calibration curves 361 using linear regression analysis. Depending on the sensitivity reached by the 362 chromatographic system employed, each analyte presents different linear responses. Both methods gave correlation coefficients (r^2) higher than 0.999 for all analytes (Table 363 364 2a). A six point calibration curve for each compound was daily performed and the 365 possible fluctuation in signal intensity was checked by injecting a standard solution at 366 two concentration levels after each eight injections.

Accuracy of the method was estimated from recovery experiments of the target
 analytes at different concentration levels. Four fortification levels in river water were
 tested for both methods; however fast-HPLC method could not be validated to the 0.75

 μ g/L level due to its lower sensitivity comparatively to HPLC method, but it was 371 validated to the other fortification levels (2.50, 3.75 and 5.00 μ g/L) as is shown in Table 372 2b. The recoveries obtained were satisfactory for all the compounds (recovery from 60 373 to 120%, with RSD values lower than 6%), except for *p*-aminophenol. In this case, a 374 value lower than 60% for 0.75 μ g/L spiking level was obtained, which could be related 375 with losses during the pre-concentration step due to the physico-chemical properties of 376 the compound.

The precision of the method was evaluated in terms of repeatability (intra-day) and
reproducibility (inter-day), exhibiting RSD values below to 3% and 6%, respectively,
for both methods (Table 2a).

The influence of the matrix in the UV signal was evaluated by comparing a calibration curve in river extract with one achieved for the standards prepared in a mixture of methanol-water (10:90, v/v). The calibration curves obtained for all the compounds exhibited similar slopes and higher Y-intercept values for the river extract, indicating that there is a slightly matrix effect, however this is similar within the linear range. In order to correct the matrix effect, the standard addition method was used.

3.4. Application to real samples

In order to demonstrate the applicability of the developed method, samples from seven rivers located in the north of Portugal were analyzed. As SPE-HPLC/DAD showed to be more sensitive, with lower MDL and MQL for all the compounds, samples were only analyzed using this method. Quantification of the river samples was done using the standard addition method and the confirmation of the positive findings was carried out by comparing the UV spectra of the peaks present in the samples with that one obtained in the standard solution chromatogram. An example of a

-	395	chromatogram of a river sample obtained with the developed methodology is shown in
1 2 3	396	Figure 4.
4 5	397	
6 7 8	398	Insert Figure 4
9 10	399	
11 12 13	400	All the compounds were detected in Portuguese rivers (Table 3). As expected,
14 15	401	paracetamol reported the lowest concentrations (up to 0.25 μ g/L), since it is efficiently
16 17	402	removed (approximately 99%) in WWTPs [2]. The results are in agreement with
18 19 20	403	concentrations of paracetamol reported in literature in rivers from Spain [17], Serbia
21 22	404	[16] and United Kingdom [9].
23 24 25	405	
26 27	406	Insert Table 3
28 29 30	407	
31 32	408	On the other hand, the principal metabolite of paracetamol (paracetamol-
33 34 35	409	glucuronide) and its main transformation product (p-aminophenol) could be detected in
36 37	410	Portuguese rivers at levels up to 3.57 μ g/L and 1.63 μ g/L, respectively, usually having
38 39 40	411	higher concentrations than paracetamol (Table 3). Given that paracetamol-glucuronide
41 42	412	entry in the environment through human excretion and it was previously detected in a
43 44 45	413	WWTP effluent at levels up to 462 μ g/L [21], WWTP effluents may be pointed out as
46 47	414	the principal source of entrance of this metabolite into surface waters.
48 49	415	<i>p</i> -Aminophenol was identified as a transformation product of paracetamol and its
50 51 52	416	presence reported in wastewaters [23]. Our results showed that <i>p</i> -aminophenol could
53 54	417	also be found in surface waters, being detected in five out of seven studied rivers, with
55 56 57	418	concentrations from below MQL (<0.23 $\mu g/L$) to 1.63 $\mu g/L$ (Table 4). Although these
58 59	419	concentrations are lower than those one reported in two Romanian rivers [40].
60 61 62 63 64		17

420	
421	
422	4. Conclusions
423	A rapid, simple and sensitive method for the simultaneous determination of
424	paracetamol, paracetamol-glucuronide and <i>p</i> -aminophenol in river water was developed.
425	The extraction procedure allowed the analysis of three related compounds with different
426	physico-chemical properties in a single step using two Oasis WAX cartridges in tandem
427	and a sample pH of 7. Recoveries higher than 60% were obtained for all the
428	compounds. The SPE procedure allows the elimination of interferences and, at the same
429	time, a pre-concentration of the analytes, depicted in MDLs from 40 to 67 ng/L.
430	Comparatively to HPLC-MS/MS, HPLC-DAD has the advantageous of being an
431	inexpensive analytical technique that can be seen as an affordable, useful and cost-
432	effective alternative for routine analysis of pharmaceuticals, their metabolites and
433	transformation products in environmental waters.
434	The developed method was applied to the determination of the selected analytes
435	in samples from seven Portuguese rivers, showing their occurrence in surface waters
436	with levels up to few micrograms per litre. The results obtained shows that besides
437	pharmaceuticals, scientific community should also focus its attention in the evaluation
438	of the presence of metabolites and transformation products in the environment, seeing
439	that, sometimes, they may be at higher concentrations than the parent compounds. The
440	proposed methodology may be applied to monitoring the behaviour of the selected
441	analytes during wastewater treatment as well as their ability to adsorb to soil and
442	sediments.
443	
444	
	18

445	Acknowledgments
446	This work has been supported by Fundação para a Ciência e a Tecnologia (FCT)
447	through the grant no. Pest-C/EQB/LA0006/2011, project PTDC/ECM/103141/2008 and
448	project PTDC/AAC-AMB/120889/2010. Lúcia H.M.L.M. Santos thanks to FCT and
449	FSE/POPH for her PhD grant (SFRH/BD/48168/2008).
450	
451	
452	References
453	[1] A. Bertolini, A. Ferrari, A. Ottani, S. Guerzoni, R. Tacchi, S. Leone, CNS Drug
454	Rev. 12 (2006) 250.
455	[2] A. Ziylan, N.H. Ince, J. Hazard. Mat. 187 (2011) 24.
456	[3] M.J.M. Bueno, M.J. Gomez, S. Herrera, M.D. Hernando, A. Agüera, A.R.
457	Fernández-Alba, Environ. Pollut. 164 (2012) 267.
458	[4] S.K. Behera, H.W. Kim, JE. Oh, HS. Park, Sci. Total Environ. 409 (2011) 4351.
459	[5] G.L. Brun, M. Bernier, R. Losier, K. Doe, P. Jackman, HB. Lee, Environ. Toxicol.
460	Chem. 25 (2006) 2163.
461	[6] M.J. Gómez, M.J.M. Bueno, S. Lacorte, A.R. Fernández-Alba, A. Agüera,
462	Chemosphere 66 (2007) 993.
463	[7] E. Gracia-Lor, J.V. Sancho, F. Hernández, J. Chromatogr. A 1218 (2011) 2264.
464	[8] Y. Vystavna, F. Huneau, V. Grynenko, Y. Vergeles, H. Celle-Jeanton, N. Tapie, H.
465	Budzinski, P. Le Coustumer, Water Air Soil Pollut. 223 (2012) 2111.
466	[9] J.P. Bound, N. Voulvoulis, Water Res. 40 (2006) 2885.
467	[10] S. Wu, L. Zhang, J. Chen, Appl. Microbiol. Biotechnol. 96 (2012).
468	[11] M.S. Kostich, J.M. Lazorchak, Sci. Total Environ. 389 (2008) 329.
469	[12] Y. Kim, K. Choi, J. Jung, S. Park, PG. Kim, J. Park, Environ. Int. 33 (2007) 370.
	19

- 470 [13] K.P. Henschel, A. Wenzel, M. Diedrich, A. Fliedner, Regul. Toxicol. Pharm. 25
 471 (1997) 220.
 472 [14] Z.H. Li, Randak, T., Vet. Med.-Czech 52 (2009) 295.
 473 [15] P.H. Roberts, K.V. Thomas, Sci. Total Environ. 356 (2006) 143.
 - 474 [16] S. Grujic, T. Vasiljevic, M. Lausevic, J. Chromatogr. A 1216 (2009) 4989.
 - 475 [17] M. Gros, S. Rodríguez-Mozaz, D. Barceló, J. Chromatogr. A 1248 (2012) 104.
 - 476 [18] S. Heitmeier, G. Blaschke, J. Chromatogr. B 721 (1999) 93.
 - 477 [19] L.S. Jensen, J. Valentine, R.W. Milne, A.M. Evans, J. Pharm. Biomed. Anal. 34
 - 478 (2004) 585.
 - 479 [20] I. Baranowska, A. Wilczek, Anal. Sci. 25 (2009) 769.
 - 480 [21] M. Sunkara, M.J.M. Wells, Environ. Chem. 7 (2010) 111.
 - 481 [22] M. Bedner, W.A. Maccrehan, Environ. Sci. Technol. 40 (2006) 516.
 - 482 [23] M. del Mar Gomez-Ramos, A. Perez-Parada, J.F. Garcia-Reyes, A.R. Fernandez-
 - 483 Alba, A. Agueera, J. Chromatogr. A 1218 (2011) 8002.
 - 484 [24] G.D. Rodrigues, L.R. Lemos, P.R. Patrício, L.H.M. Silva, M.C.H. Silva, J. Hazard.
 485 Mat. 192 (2011) 292.
 - 486 [25] W.W. Buchberger, J. Chromatogr. A 1218 (2011) 603.
 - 487 [26] P. Paíga, L.H.M.L.M. Santos, C.G. Amorim, A.N. Araújo, M.C.B.S.M.
 - 488 Montenegro, A. Pena, C. Delerue-Matos, Environ. Sci. Pollut. Res. 20 (2013) 2410.
 - 489 [27] J.L. Santos, I. Aparicio, E. Alonso, M. Callejon, Anal. Chim. Acta 550 (2005) 116.
 - 490 [28] S. He, Q. Chen, Y. Sun, Y. Zhu, L. Luo, J. Li, Y. Cao, J. Chromatogr. B 879
 491 (2011) 901.
 - 492 [29] R. Cueva-Mestanza, Z. Sosa-Ferrera, M.E. Torres-Padron, J.J. Santana-Rodriguez,
 - 493 J. Chromatogr. B 863 (2008) 150.

1	494	[30] M.J. Gómez, A. Agüera, M. Mezcua, J. Hurtado, F. Mocholí, A.R. Fernández-
1 2 3	495	Alba, Talanta 73 (2007) 314.
2 3 4 5 6 7	496	[31] E. Gracia-Lor, M. Martínez, J.V. Sancho, G. Peñuela, F. Hernández, Talanta 99
6 7 8	497	(2012) 1011.
9 0	498	[32] J. Martín, W. Buchberger, J. Luis Santos, E. Alonso, I. Aparicio, J. Chromatogr. B
1 2 3	499	895 (2012) 94.
3 4 5	500	[33] B. Kasprzyk-Hordern, R.M. Dinsdale, A.J. Guwy, Anal. Bioanal. Chem. 391
6 7 8	501	(2008) 1293.
9 0	502	[34] M. Laven, T. Alsberg, Y. Yu, M. Adolfsson-Erici, H.W. Sun, J. Chromatogr. A
1 2 3	503	1216 (2009) 49.
3 4 5	504	[35] USEPA, in, U.S. Environmental Protection Agency Office of Water Office of
6 7	505	Science and Technology Engineering and Analysis Division Washington, DC, 2007.
8 9 0	506	[36] N. Fontanals, R.M. Marce, F. Borrull, J. Chromatogr. A 1152 (2007) 14.
1 2 3	507	[37] N. Fontanals, P.A.G. Cormack, R.M. Marce, F. Borrull, TrAC Trend Anal. Chem.
3 4 5	508	29 (2010) 765.
6 7	509	[38] www.chemicalize.org (accessed in 29 th October 2012)
8 9 0	510	[39] www.chemspider.org (accessed in 29 th October 2012)
0 1 2	511	[40] I. Badea, L. Axinte, L. Vladescu, Environ. Monit. Assess. 185 (2013) 2367.
3 4 5	512 513	
6 7 8	514	
9 0		
1 2		
3 4		
5 6		
7		
8 9		
0		
1 2		21
3 4		


-	515	Figure captions:
1 2 3	516	
4 5	517	Figure 1 – Recoveries obtained for the target analytes in ultra-pure water with
6 7 8	518	different SPE cartridges. PCT-G – Paracetamol-glucuronide; PAP – <i>p</i> -Aminophenol;
9 10	519	PCT – Paracetamol
11 12 13	520	
14 15	521	Figure 2 – Recoveries obtained in ultra-pure water at different sample's pH. a)
16 17 18	522	Paracetamol-glucuronide; b) <i>p</i> -Aminophenol; c) Paracetamol
19 20	523	
21 22	524	Figure 3 – Example of a chromatogram of a standard mixture 10 μ g/mL for the
23 24 25	525	selected analytes analyzed by a) HPLC and b) Fast-HPLC. 1 – Paracetamol-
26 27	526	glucuronide; 2 – <i>p</i> -Aminophenol; 3 – Paracetamol
28 29 30	527	
31 32	528	Figure 4 – Chromatogram of Douro river sample. 1 – Paracetamol-glucuronide; 3 –
33 34 35	529	Paracetamol
	530 531 532 533	

Highlights


- Determination of paracetamol, paracetamol-glucuronide and *p*-aminophenol in rivers
- Development of a SPE procedure for their simultaneous determination
- All compounds were detected in rivers with concentrations up to few $\mu g/L$
- First time that paracetamol-glucuronide was found in river water

A Cepter Manus

Figure 1



X







Figure 3



X

Figure 4






Table 1

ACCEPTED MANUSCRIPT

G

Table 1 – Physico-chemical properties of the selected compounds

	Chemical structure	Formula	CAS no.	M _w	Solubility (g L ⁻¹)	pKa ^a	$\log K_{OW}^{b}$
Paracetamol (PCT)	HO HO CH ₃	C ₈ H ₉ NO ₂	130-90-2	151.16	12.8	9.46	0.46
Paracetamol-glucuronide (PCT-G)		C ₁₄ H ₁₇ NO ₈	120595-80-4	327.29	27.7	3.17; 12.22	-1.23
p-Aminophenol (PAP)	H ₂ N OH	C ₆ H ₇ NO	123-30-8	109.13	15.0	5.43; 10.40	0.24
^a Adapted from refere ^b Adapted from refere							

Table 2 – **a**) Linearity, detection and quantification limits of the method (MDL, MQL), precision intra- and inter-day of the developed SPE-HPLC/DAD and SPE-fast-HPLC/DAD methods. **b**) Recoveries obtained, expressed in percentage (%), for the selected analytes in river water for the developed methods.

a)													
SPE-HPLC/DAD								SPE-fast-HPLC/DAD					
Compound	Linear range (µg/L)	Correlation coefficient (r^2)	MDL (µg/L)	MQL (µg/L)	Precision intra-day (% RSD)	Precision inter-day (% RSD)	Linear range (µg/L)	Correlation coefficient (r^2)	MDL (µg/L)	MQL (µg/L)	Precision intra-day (% RSD)	Precision inter-day (% RSD)	
PCT-G	50-1500	0.9999	0.040	0.134	0.547	0.836	50-1500	0.9997	0.199	0.664	2.75	5.06	
PAP	50-1500	0.9999	0.067	0.225	0.368	0.632	500-1250	0.9990	0.275	0.905	2.14	4.39	
PCT	50-1500	0.9999	0.042	0.141	0.097	0.958	50-1500	0.9996	0.210	0.701	0.913	3.56	

b)

		SPE-HP	LC/DAD		SF	PE-fast-HPLC/DA	AD
-			b)				
Compound		Spiking le	vel (µg/L)	Spiking level (µg/L)			
Compound -	0.75	2.50	3.75	5.00	2.50	3.75	5.00
PCT-G	100 (2.4)	117 (2.1)	110 (0.7)	112 (3.0)	108 (2.5)	99 (0.7)	112 (1.2)
PAP	53 (5.5)	62 (4.3)	64 (0.5)	60 (0.2)	62 (3.7)	65 (4.0)	59 (2.7)
PCT	99 (1.5)	103 (2.3)	107 (3.9)	109 (0.02)	106 (3.5)	105 (4.0)	108 (0.9)

Sample	Paracetamol-glucuronide	p-Aminophenol	Paracetamol
Cabrum river	<mql< td=""><td>n.d.</td><td><mdl< td=""></mdl<></td></mql<>	n.d.	<mdl< td=""></mdl<>
Douro river	<mql< td=""><td>n.d.</td><td><mql< td=""></mql<></td></mql<>	n.d.	<mql< td=""></mql<>
Ave river	0.36 (±0.02)	1.63 (±0.05)	0.17 (±0.002)
Leça river	3.57 (±0.06)	1.25 (±0.02)	0.25 (±0.01)
Tâmega river	<mql< td=""><td>0.40 (±0.01)</td><td><mdl< td=""></mdl<></td></mql<>	0.40 (±0.01)	<mdl< td=""></mdl<>
Lima river	<mql< td=""><td><mql< td=""><td><mdl< td=""></mdl<></td></mql<></td></mql<>	<mql< td=""><td><mdl< td=""></mdl<></td></mql<>	<mdl< td=""></mdl<>
Minho river	0.18 (±0.01)	0.52 (±0.05)	<mdl< td=""></mdl<>

Table 3 – Concentration of the analyzed compounds, expressed in $\mu g/L$, in river waters from the north of Portugal. Standard deviation (SD) values are indicated in brackets.

n.d. – not detected; <MDL – below method detection limit; <MQL – below method quantification limit

۲ ر ر